

Certificate of Compliance

Certificate Number 20080711 – SA5219F
Report Reference SA5219, 1999 June 21
Issue Date 2009 July 21

Page 1 of 1



Issued to:

BEACON MEDICAL PRODUCTS LLC, DBA BEACONMEDAES

1800 OVERVIEW DR
ROCK HILL, SC 29730 USA
Telephone: 803-817-5600
Fax: 803-817-5650

*This is to certify that
representative samples of*

Station Inlets and Outlets

Station outlets for medical gas.

Gas Service	Model
Oxygen	DN600, DN700, DN800, DN900
Nitrous Oxide	DN610, DN710, DN810, DN910
Vacuum	CN620, CN720, CN820, CN920
Air	DN630, DN730, DN830, DN930
Carbon Dioxide	DN640, DN740, DN940
Nitrogen	DN750
WAGD	CN660, CN760, CN860, CN960
O ₂ /CO ₂	DN770, DN970
Instrument Air	DN780

Have been investigated by Underwriters Laboratories Inc.® in accordance with the Standard(s) indicated on this Certificate.


Standard(s) for Safety:

The basic standard used to investigate products in this category is UL 1331, "Station Inlets and Outlets."

Additional Information:

None.

Only those products bearing the UL Listing Mark for the US and Canada should be considered as being covered by UL's Listing and Follow-Up Service meeting the appropriate requirements for US and Canada.

The UL Listing Mark for the US and Canada generally includes: the UL in a circle symbol with "C" and "US" identifiers:  the word "LISTED"; a control number (may be alphanumeric) assigned by UL; and the product category name (product identifier) as indicated in the appropriate UL Directory.

Look for the UL Listing Mark on the product

Issued by:

Tiffany Ladage

Reviewed by:

Robert Zeman

Tiffany Ladage, Customer Service Specialist I
Underwriters Laboratories Inc.

Robert Zeman, Section Manager
Underwriters Laboratories Inc.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL

For questions in The United States of America you may call 1-877-UL-HELPS.