


INSTRUMENTATION DETAILS			
SENSOR	DESCRIPTION	SETTING	ACTION
S1	BACKUP VAC SWT.	15" Hg ON THE FALL	BACKUP OPERATION
S2	VACUUM TRANSDUCER	VAC-15PSIG	OPERATION
S3	HIGH DISCH TEMP	5HP - 250°F 6.4-15HP - 350°F	SHUTDOWN



Revised	Drawn	12/19/13	MDR	03/15/10
This drawing and the information contained herein remain the property of BeaconMedias and may not be used for other than the purposes for which it is loaned without the expressed written permission from BeaconMedias, Inc.				Scale: NTS
PNEUMATIC DIAGRAM TX-EXP-QX 5-15HP CLAW VAC SPC		Part Number: 4107 8517 00 DNE HOP131370 Rev: 01 Sheet 1 of 1 DO NOT SCALE THIS DOCUMENT		