


INSTRUMENTATION DETAILS			
SENSOR	DESCRIPTION	SETTING	ACTION
S1	BACKUP VAC. SWT.	15" Hg ON THE FALL	BACKUP OPERATION
S2	VACUUM TRANSDUCER	VAC-15PSIG	OPERATION
S3	HIGH DISCH TEMP	15HP - 350°F	SHUTDOWN



Rev: 00	Date: 12/18/13	Drawn: JWO	Scale: NTS
		This drawing and the information contained therein remain the property of BeaconMedics and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedics Engineering.	
Description: PNEUMATIC DIAGRAM PX 15HP CLAW VAC SPC		Part Number: 4107 8550 17 Rev: 00 HOP131370	
Sheet 1 of 1		DO NOT SCALE THIS DOCUMENT	