This drawing and the information contained thereon remain the property of BeaconMedæs and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedæs Engineering.

Notes:

1. DISCHARGE MUFFLERS WILL BE SHIPPED LOOSE.
2. DX MODULE W/ PANEL WEIGHT: 1,847 LBS
   DX MODULE WEIGHT: 1,671 LBS
   240G TANK WEIGHT: 713 LBS
3. ☒ INDICATES LIFTING POINT.
4. ⊕ DENOTES CENTER OF GRAVITY.

SYSTEM CONNECTIONS	
ID	DESCRIPTION
I	INLET LOCATION
D	DISCHARGE LOCATION
C	CONDENSATE DRAIN LOCATION
T	TANK DRAIN





240 GAL VACUUM RECEIVER
3" FLANGED CONNECTIONS
(SCALE 1:11)











INSPECTION CHECKLIST
UNLESS OTHERWISE SPECIFIED:

ALL ITEMS MARKED BY ONE OF THE ABOVE SYMBOLS ARE TO BE INSPECTED

EQUIPMENT TO BE USED DURING RECEIVING INSPECTION SHALL BE PER QUALITY SYSTEM PROCEDURES

Revised:	Date:	Drawn: MDB	Date: 4/15/2010

BEACONMEDÆS

UNLESS OTHERWISE SPECIFIED ALL DIMENSIONS ARE IN INCHES AND TOLERANCES ARE:
XX ± .02
XXX ± .005
XXXX ±.0005
ANGLES ± 2°

Scale: 1/11

Size: D

Description: GENERAL ASSEMBLY 7.5HP QX CLAW VAC MOD 240G RECEIVER STANDARD AND O2 ASSURED

Part Number: 4107 8513 39

DN: USN100174

Rev: 00

Sheet 1 of 1

DO NOT SCALE THIS DOCUMENT