


INSTRUMENTATION DETAILS			
SENSOR	DESCRIPTION	SETTING	ACTION
S1	BACKUP VAC SWT.	15" Hg ON THE FALL	BACKUP OPERATION
S2	VACUUM TRANSDUCER	VAC-15PSIG	OPERATION
S3	HIGH DISCH TEMP	10-15HP - 350°F	SHUTDOWN



Revised	JWD	Date	01/28/14	Drawn	MDB	Date	05/16/12
This drawing and the information contained herein remain the property of BeaconMedias and may not be used for other than the purposes for which it is loaned without the expressed written permission from BeaconMedias, Inc.							
Description <b>PNEUMATIC DIAGRAM TX          10-15HP CLAW VAC MOD</b>						Part Number <b>4107 8539 65</b>	
DNE HOP131370						Rev 01	
Sheet 1 of 1						DO NOT SCALE THIS DOCUMENT	