


INSTRUMENTATION DETAILS			
SENSOR	DESCRIPTION	SETTING	ACTION
S1	BACKUP VAC SWT.	15" Hg ON THE FALL	BACKUP OPERATION
S2	VACUUM TRANSDUCER	VAC-15PSIG	OPERATION
S3	HIGH DISCH TEMP	5-10HP - 250°F	SHUTDOWN



Revised	Date	Drawn	JEB	Date	12/01/10
		<small>This drawing and the information contained herein remain the property of Beacon Medias and may not be used for other than the purpose for which it is issued without the expressed written permission from Beacon Medias Engineers.</small>			
Description		Part Number			
PNEUMATIC DIAGRAM QX		4107 8523 39			
5-10HP LUBE VAC MOD		USN100815		00	
Sheet 1 of 1		DO NOT SCALE THIS DOCUMENT			