ALL INTERCONNECTING PIPE AND FITTINGS SHALL BE INSTALLED BY THE CONTRACTOR. D 2. ALL PIPING SHALL BE INSTALLED IN ACCORDANCE WITH NFPA 99. PIPING LAYOUT IS DIAGRAMMATIC ONLY MAIN VACUUM GAGE--MAIN LINE VACUUM SWITCH FINAL ARRANGEMENT WILL BE SITE SPECIFIC. CONTRACTOR WILL BE RESPONSIBLE FOR FIELD VERIFICATION AND COORDINATING ACTUAL LOCATION WITH ALL OTHER TRADES. SOURCE SHUTOFF VALVE-3. CONTRACTOR SHALL VERIFY MOTOR VOLTAGE, PHASE AND AMP RATINGS BEFORE STARTING ELECTRICAL INSTALLATION; DISCHARGE 3/4" NPT AND MAKE CERTAIN THE VOLTAGE SUPPLIED BY THE HOSPITAL IS THE SAME. FOR 1.5 HP AND 2 HP -FROM HOSPITAL 4. THE ROOM REQUIRES DEDICATED MECHANICAL VENTILATION WITH AN ALLOWABLE TEMPERATURE VARIANCE OF 40° TO ALARM-105°F. (FOR HIGHER AMBIENTS, CONSULT FACTORY) WIRING ACCESS 5. ELECTRICAL POWER SHOULD BE SUPPLIED FROM THE EQUIPMENT SYSTEM BRANCH OF THE ESSENTIAL ELECTRICAL FOR INTAKE PIPE SIZING REFER TO STANDARD SYSTEM (EMERGENCY POWER). PRESSURE DROP TABLES. TOTAL PRESSURE DROP SHOULD BE LESS THAN 2" Hg AT A 6. EQUIPMENT, INSTALLATION, AND TESTING OF THE MEDICAL VACUUM SYSTEM SHALL COMPLY WITH NFPA 99 SYSTEM VACUUM LEVEL OF 15" Hg. AND ALL STATE AND LOCAL CODES OR ORDINANCES. 7. WIRE LAG PUMP RUNNING ALARM REMOTE CONTACTS IN THIS CONTROL SYSTEM TO MASTER ALARM SYSTEM AS REQUIRED BY NFPA 99. SINGLE POINT POWER-WIRING ACCESS 8. ALL ACCESSORIES, PIPE AND FITTINGS EXCEPT DISCHARGE FLEX HOSE BEYOND INLET AND DISCHARGE PORTS ARE -Ø0.75 MTG HOLE (4 PLACES) SUPPLIED AND INSTALLED BY OTHERS. 9. WHEN DETERMINING THE TOTAL PIPE LENGTH, ADD ALL THE STRAIGHT LENGTHS OF PIPE TOGETHER IN ADDITION TO THE NUMBER OF ELBOWS TIME THE EFFECTIVE PIPE LENGTH FOR THAT PIPE SIZE. (SEE EQUIVALENT PIPE LENGTH TABLE & EXAMPLE CALCULATION IN THE O&M MANUAL) PIPE EXHAUST OUTSIDE, TURN DOWN AND SCREEN PER NFPA 99 -DISCHARGE FLEX HOSE (1 REQUIRED AND ARE SUPPLIED WITH UNIT) DRIP LEG VALVE SUPPLIED AND INSTALLED BY OTHERS (1 REQUIRED) FIELD PIPING INLET ===== FIELD PIPING OUTLET **HOSPITAL** RECEIVER MANUAL DRAIN INLET CONNECTION 1/2" NPT, CONTRACTOR 1-1/2" NPT SHOULD PIPE TO NEAREST EXHAUST PIPE SIZE TABLE (SEE NOTE 9) FLOOR DRAIN PIPE LENGTH **UNIT SIZE** 0' - 250' 251' - 500' SPX 1.5 HP 1.00" NPT 1.25" NPT PIPE LENGTH **UNIT SIZE** 0' - 125' 126' - 500' NO FOUNDATION OR INERTIA PAD IS REQUIRED -NO VIBRATION PADS ARE SPX 2 HP 1.00" NPT 1.25" NPT (HOUSEKEEPING PAD IS OPTIONAL) NECESSARY AT THIS LOCATION EFFECTIVE PIPE LENGTH TABLE FOR ELBOWS 1.0" NPT 1.25" NPT EFFECTIVE PIPE LENGTH (FT) 3 3.4 INSTALLATION DIAGRAM SX 1.5-2HP DRY VTM V 30G BASIC (2 HP WITH 30 GALLON RECEIVER SHOWN) Sheet 1 of 1